

FavorPrep™ Viral Nucleic Acid Extraction Kit II

- For isolation of DNA from cell-free fluid such as, serum, plasma, body fluid and cell cultured supernatant

Cat.No. : FVNKK 000-2, 4 Preps
FAVVK 002, 50 Preps
FAVVK 002-1, 100 Preps
FAVVK 002-2, 300 Preps

Kit Contents:

(For Research Use Only)

	FAVVK 000-2 (4 preps_sample)	FAVVK 002 (50 preps)	FAVVK 002-1 (100 preps)	FAVVK 002-2 (300 preps)
AD Buffer * (concentrate)	0.4 ml	4 ml	8 ml	24 ml
VNE Buffer	1.8 ml x 2	30 ml	60 ml	180 ml
Wash Buffer 1 * (concentrate)	0.9 ml x 2	22 ml	44 ml	132 ml
Wash Buffer 2 * (concentrate)	1.5 ml	20 ml	20 ml x 2	50 ml x 2
RNase-free Water	0.5 ml	6 ml	12 ml	30 ml
VNE Column	4 pcs	50 pcs	100 pcs	300 pcs
Collection Tube	8 pcs	100 pcs	200 pcs	600 pcs
Elution Tube	4 pcs	50 pcs	100 pcs	300 pcs
User Manual	1	1	1	1

* Preparation of AD Buffer, Wash Buffer W1 and Wash Buffer W2 for first use:

Cat. No:	FAVVK000-2 (4 preps)	FAVVK002 (50 preps)	FAVVK002-1 (100 preps)	FAVVK 002-2 (300 preps)
ethanol volume for AD Buffer	3 ml	30 ml	60 ml	180 ml
ethanol volume for Wash Buffer 1	0.33 ml	8 ml	16 ml	48 ml
ethanol volume for Wash Buffer 2	6 ml	80 ml	80 ml	200 ml

Specification:

Principle: spin column (silica membrane)

Sample: 150 µl cell-free fluid such as serum, plasma, body fluid and cell cultured supernatant

Fragment size: 100 bp ~30 kb

Recovery rate: 70 ~ 90 %


Binding capacity: 30 µg

Elution volume: 40 ~ 50 µl

Operation time: 20 min

Important Notes:

1. Make sure everything is RNase-free when handling this system.
2. Buffers provided in this system contain irritants. Wear gloves and lab coat when handling these buffers.
3. Add required ethanol (96-100%) to AD Buffer, Wash Buffer 1 and Wash Buffer 2 before use.
4. Preheat RNase-free water to 70°C for elution step. (step:10)



General Protocol:

Please Read Important Notes Before Starting Following Steps.

HINT: Preheat RNase-free water 70 °C for step 11 (elution step).

1. Transfer 200 µl of sample (serum, plasma, body fluids or cell cultured supernatant) into a microcentrifuge tube (not provided).
--If prepared sample is less than 200 µl , adjust sample volume to 200µl with PBS (not provided).
2. Add 500 µl of VNE Buffer the sample, mix well by vortexing, and incubate for 10 minutes at room temperature.
3. Add 550 µl of AD Buffer (ethanol added) to the sample mixture and mix well immediately by plus-vortexing.
--Make sure that ethanol has been added into AD Buffer when first open.
4. Combine a VNE column with a Collection Tube (provided).
5. Transfer up to 750 µl of sample mixture (ethanol added) to the VNE Column, centrifuge at 8,000 x g for 1 min then discard the flow-through. Combine the VNE Column with the used Collection Tube.
6. Transfer the rest of sample mixture (ethanol added) to the VNE Column, centrifuge at 8,000 x g for 1 min.
Discard the flow-through and the Collection Tube. Combine the VNE Column with a new Collection Tube (provided).
7. Add 500 µl of Wash Buffer 1 (ethanol added) to the VNE Column, centrifuge at 8,000 x g for 1 min then discard the flow-through.
Combine the VNE Column with the used Collection Tube.
--Make sure that ethanol (96~100%) has been added into Wash Buffer 1 when first open.
8. Add 750 µl of Wash Buffer 2 (ethanol added) to the VNE Column, centrifuge at 8,000 x g for 1 min then discard the flow-through.
Combine the VNE Column with the used Collection Tube.
--Make sure that ethanol (96~100%) has been added into Wash Buffer 2 when first open.
9. Repeat step 8. Add 750 µl of Wash Buffer 2 (ethanol added) to VNE Column, centrifuge at 8,000 x g for 1 min then discard the flow-through. Combine the VNE Column with the used Collection Tube.
10. Centrifuge at full speed (~18,000 X g) for an additional 3 min to dry the VNE column. Discard the flow-through and the Collection Tube.
--Important step! This step will avoid the residual liquid to inhibit the subsequent enzymatic reactions.
11. Combine the VNE Column with a Elution Tube (provided). Add 50 µl of preheated RNase-free Water to the membrane center of the VNE Column. Stand VNE Column for 2 min.
--Important step! For effective elution, make sure that the RNase-free Water is dispensed onto the membrane center and is absorbed completely.
12. Centrifuge for 2 min to elute the nucleic acid.
12. Store nucleic acid at -70 °C.

Troubleshooting

Problems	Possible reasons	Solutions
Low or no yield of genomic DNA		
	Incorrect preparation of Wash Buffer 1 or Wash Buffer 2	
	AD Buffer, Wash Buffer W1 and Wash Buffer W2 is not mixed with ethanol before use	Make sure that the correct volumes of ethanol (96-100 %) is added into AD Buffer, Wash Buffer W1 and Wash Buffer W2 when first open. Repeat the extraction procedure with a new sample.
	The volume or the percentage of ethanol is not correct before adding into AD Buffer, Wash Buffer W1 and Wash Buffer W2	Make sure that the correct volumes of ethanol (96- 100 %) is added into AD Buffer, Wash Buffer W1 and Wash Buffer W2 when first use. Repeat the extraction procedure with a new sample.
	Elution of genomic DNA is not efficient	
	RNase-free water not completely absorbed by column membrane	After RNase-free water is added, stand the VNE Column for 2 min before centrifugation.
Column is clogged		
	Sample is too viscous	Reduce the sample volume.
Degradation of elutated DNA		
	Sample is old	Always use fresh or well-stored sample viral nucleic acid extraction.